

Webtec Hydraulic Flow & Pressure Calibration Service

**Reliable, repeatable measurements every
time from your instruments.
Total confidence in quantity measurements.**



Milwaukee, WI 53235, USA

Tel: +1 (414) 769-6400
sales-us@webtec.com

St. Ives, Cambs. PE27 3LZ, UK

Tel: +44 (0) 1480 397 400
sales-uk@webtec.com

www.webtec.com

Features

- Calibrations traceable to national and international standards.
- Robust monitoring system ensures consistent compliance.
- All instrumentation employed has full traceability.
- Calibration certificate demonstrating traceability; suitable for ISO9001 compliance.
- Calibration lead time is typically less than 2 weeks.




Certificate No.8242

FLOWCAL-BU-ENG-3568.pdf 06/17
(Issue 2)

Hydraulic measurement and control

Flow Calibration Method

Below is the simplified hydraulic circuit employed in the calibration of all flow meters. It is a ratio-metric system whereby the oil is pumped in series through a reference meter and a Unit Under Test (UUT). Comparison of the reference and UUT flow shows the deviation and the UUT calibration factor can be adjusted accordingly.



Typical Calibration

Flow meters are checked at various points throughout their range to ensure they meet the exacting criteria. If any of the flow points fail to meet the accuracy specification repairs or adjustments are made.

Reference Meter Traceability

Reference meters are calibrated in a laboratory that maintains traceability to national and international standards, PTB (Germany), LNE (France), UKAS (UK) and NIST (USA).

The laboratory maintains an overall measurement uncertainty of 0.03% and has ISO9001 accreditation. It also complies with the ANSI/NCSL Z540-1-1994 laboratory standards. The laboratory operates within a Quality Management System according to ISO 9001:2008 and ISO/IEC 17025:2005.

Webtec reference meters are calibrated annually and are maintained using a rolling system whereby meters are calibrated just before and just after their installation anniversary. This ensures that all calibrations are conducted against a 'known good reference'.

Calibration uncertainty around our reference meters is <0.2% and all meters produce stable results within this limit. During reference meter calibration viscosity variance is kept <1cSt so as to have negligible effect.

Flow Calibration Test Equipment

Webtec use a C2000 data acquisition system with a customised interface unit that conditions signals appropriately. The interface unit is monitored under the internal calibration system and all signal conversions are checked annually to maintain errors at better than <0.01%. Instruments used for the checks are calibrated annually by a UKAS accredited laboratory.

Oil condition is continually monitored through a particle counter to ensure it meets good cleanliness standards (NAS 8 or better). The oil is also sent for analysis annually to provide external verification.

Temperature sensors engaged to measure the test fluid temperature are monitored under the internal calibration system and are checked annually to maintain accuracy at better than $\pm 1^{\circ}\text{C}$. Instruments used for these checks are calibrated annually by a UKAS accredited laboratory.

Fluid Viscosity

Flow meters are typically calibrated at a mean viscosity of 21cSt which is equivalent to ISO32 oil at 50°C. This temperature is accurately maintained to ensure a viscosity variance of <5cSt during any calibration. Calibration at other viscosities is possible, please contact sales.

Flow Calibration Certificate

WEBTEC
 Webtec Products Limited
 Nuffield House
 95, Innes
 Cambridge Road
 PE27 3LJ, UK
 www.webtec.com
 Registered in England
 England & Wales

Certificate of Flow Calibration

Date: 08 February 2017
 Certificate No: 161553-LP00060
 Part No: FT9503-24
 Model No: CT300-CAN-B18-B
 Serial No: LP00060
 Operator: CS

Maximum allowable error ± 1% indicated reading over top 80% of range

Reference Flow (LPM)	Device Under Test Output (LPM)	Device Under Test Flow (LPM)	Flow Error (LPM)
238.73	N/A	200.00	1.30
248.96	N/A	250.00	1.04
199.20	N/A	200.00	0.80
149.35	N/A	150.00	0.61
99.59	N/A	100.00	0.12
75.29	N/A	75.00	-0.28
49.84	N/A	50.00	0.16
30.11	N/A	30.00	-0.11
15.11	N/A	15.00	-0.11
8.11	N/A	8.00	-0.11

Oil Test Conditions
 Type: Esso Nuto H32
 Temperature: 45 ± 0.5 Centigrade
 Viscosity: 21 cSt at 50° Centigrade
 Pressure: 200 ± 50 PSI

Signed By: K. Parsons
 Signature:

Webtec recommends that a hydraulic flowmeter or hydraulic tester is re-calibrated every 6000 hours or every 12 months, whichever is sooner

Hydraulic measurement and control

STATEMENT OF TRACEABILITY

All measuring methods and test equipment are traceable to National or International Standards and are calibrated in accordance with our ISO9001 Quality Procedures

Flow
 Flow is tested against the relevant reference flow clocks detailed below

Reference Type	Serial No	Flow Range (l/min)	Calibration valid until	Calibration Certificate No.
LT600 Turbine Flow Meter	LU00007	50 - 800	31/03/2017	24178-LU00007
VC3 Gear Flow Meter	F10-0684	1 - 140	31/03/2017	24178-F10-0684
VC08 Gear Flow Meter	9198	0.1 - 20	31/03/2017	24948-9198
VC12 Gear Flow Meter	0701	2 - 400	31/03/2017	24948-0701

Pressure
 The pressure reference is a dead-weight pressure tester

Reference Type	Serial No	Pressure Range (bar)	Calibration valid until	Calibration Certificate No.
Budenberg Tester	60442004	1 - 1100	13/08/2016	161576-13

Temperature
 Temperature is checked using equipment tested against a digital reference thermometer

Reference Type	Serial No	Temperature Range (°C)	Calibration valid until	Calibration Certificate No.
Fisher 110 NTC Digital Thermometer - Probe	33940310105	10 - 70	21/12/2017	00286422

Electronic Equipment
 The conductor is checked using the equipment detailed below

Reference Type	Serial No	Signals	Calibration valid until	Calibration Certificate No.
Agilent 34401A Digital Multimeter	UC46180507	Frequency & mA Input	31/03/2016	00068302

Webtec Ltd - 01152
 Date: 08 February 2017

The flow calibration certificate shows the details of the meter, a table of results and the conditions during the test. The second page displays the reference instrumentation employed during the calibration.

Pressure Calibration Method

Verification of pressure calibration is conducted on a Dead-Weight-Tester where the pressure is generated using sets of calibrated weights. These systems are extremely robust and reliable, providing total confidence in the results.

Pressure Calibrator Traceability

The Dead-Weight-Testers are calibrated at laboratories that maintain traceability to national and international standards, PTP (Germany), LNE (France), UKAS (UK) and NIST (USA).

The laboratories are accredited under their national schemes and maintain a measurement uncertainty for pressure at <0.02%.

Other Calibration Test Equipment

Some sensors require pressure calibration output signals to be recorded as an electrical value. Measurement instruments used for this purpose are calibrated to the same level of traceability as the Dead-Weight-Testers and have an accuracy of <0.06% and a level of calibration uncertainty <0.1%.

Pressure Calibration Certificate

WEBTEC
 Webtec Products Limited
 Nuffield House
 95, Innes
 Cambridge Road
 PE27 3LJ, UK
 www.webtec.com
 Registered in England
 England & Wales

Certificate of Pressure Calibration

Date: 08 February 2017
 Certificate No: 123456-1122100923
 Part No: FT10422-000
 Model No: SR-PTT-600-05-OC-CAN
 Serial No: 1122100923
 Batch No: 12345
 Operator: CS

Maximum allowable error ± 0.5% of full scale reading

Reference Pressure (bar)	Device Under Test Output (bar)	Device Under Test Pressure (bar)	% Error
0.0	N/A	0.00	0.00
150.0	N/A	149.30	-0.12
300.0	N/A	298.40	-0.10
450.0	N/A	450.00	0.00
600.0	N/A	600.50	0.08

Oil Test Conditions
 Type: Total/Anala ZR 22
 Temperature: 105 ± 0.2 Centigrade
 Viscosity: 50 cSt at 20° C

Signed By: K. Parsons
 Signature:

Webtec recommends that a Pressure Transducer or Gauge is checked for accuracy every 6000 hours or every 12 months, whichever is sooner

Hydraulic measurement and control

STATEMENT OF TRACEABILITY

All measuring methods and test equipment are traceable to National or International Standards and are calibrated in accordance with our ISO9001 Quality Procedures

Flow
 Flow is tested against the relevant reference flow clocks detailed below

Reference Type	Serial No	Flow Range (l/min)	Calibration valid until	Calibration Certificate No.
LT600 Turbine Flow Meter	LU00007	50 - 800	31/03/2017	24178-LU00007
VC3 Gear Flow Meter	F10-0684	1 - 140	31/03/2017	24178-F10-0684
VC08 Gear Flow Meter	9198	0.1 - 20	31/03/2017	24948-9198
VC12 Gear Flow Meter	0701	2 - 400	31/03/2017	24948-0701

Pressure
 The pressure reference is a dead-weight pressure tester

Reference Type	Serial No	Pressure Range (bar)	Calibration valid until	Calibration Certificate No.
Budenberg Tester	60442004	1 - 1100	13/08/2016	161576-13

Temperature
 Temperature is checked using equipment tested against a digital reference thermometer

Reference Type	Serial No	Temperature Range (°C)	Calibration valid until	Calibration Certificate No.
Fisher 110 NTC Digital Thermometer - Probe	33940310105	10 - 70	21/12/2017	00286422

Electronic Equipment
 The conductor is checked using the equipment detailed below

Reference Type	Serial No	Signals	Calibration valid until	Calibration Certificate No.
Agilent 34401A Digital Multimeter	UC46180507	Frequency & mA Input	31/03/2016	00068302

Webtec Ltd - 01152
 Date: 08 February 2017

The pressure calibration certificate records the details of the device, a table of measured results and the conditions during the test. The second page displays the reference instrumentation employed during the calibration.