



# Installation notes HPMComm 7.1.9.8 and HPM6000 firmware N102



# Release notes for HPMComm 7.1.9.8 and firmware (N102) for HPM6000

## HPM6000 series firmware update

Improvements for the HPM6000 series (N102):

- Channel name visible in the curve view
- Submission of decimal figures
- Trigger und Trigger Logic measurement with decimal figures as a start and stop value
- Reset-functions are also available in the project mode
- Project settings were still kept, after the project measurement
- Extensive project edit options

## Installation notes

Files with a \*.CAB extension are used for the update. These files are copied to the measuring instrument. A USB mass storage device should be used to transfer the files.

For updating the firmware of The HPM6000 please follow the steps below:

1. Copy all \*.CAB files which are in the downloaded file, from the PC to the root folder of a USB stick (without a sub-directory).
2. Disconnect all sensors that may be connected to your measuring instrument.
3. Start up the instrument and wait until the message "NO SENSORS" appears.
4. Connect the USB device to the instrument and wait until "USB" is displayed.
5. Briefly press the ON/OFF key.
6. Wait until the device display "Update found!"
7. Press "OK" to follow the displayed instructions. Wait until the update process has finished. This may take up to 10 minutes.
8. The firmware on the instrument is now up to date.

**Please note that due to the new features the language in the HPM6000 initially switches to English. To change to the desired Language, please use only the software HPMComm 7.1.9.8. IMPORTANT: Firmware N102 needs HPMComm 7.1.9.8 to work correctly.**

## HPMComm 7.1.9.8

Improvements for the HPMComm 7.1.9.8

- The frequency converter can now be configured with Windows 7 64 bit
- Improved documentation display for one channel only, no column mismatch
- Easier csv export
- Issue with comma and dot solved
- Printer presents can be changed and adopted
- Header, footer and text are saved together
- No column for mean value reserved in exported csv files

## Installation notes

Before installing HPMComm 7.1.9.8 please uninstall the existing HPMComm version from your PC. Than execute the .exe to install the new Version 7.1.9.8.

Please visit our website where you can download the files [www.webtec.com](http://www.webtec.com)

For Sales & Service contact  
Auskunft & Beratung  
Contact Service commercial & maintenance  
Para más información sobre ventas y servicios contactar con

Distributor - Vertriebspartner - Distributeur - Distribuidor



St. Ives, Cambs, PE27 3LZ, UK  
Tel: +44 (0) 1480 397 400 - sales-uk@webtec.com

中国:

Tel: +86-21-61498441 - sales-cn@webtec.com

France

Tel: +33 (0) 3 27 82 94 56 - ventes-fr@webtec.com

Deutschland

Tel: +49 (0)231-9759-747 - vertrieb-de@webtec.com

U.S.A & Mexico

Tel: +1-800-932-8378 - sales-us@webtec.com

**www.webtec.com**



Certificate No.8242

Webtec reserve the right to make improvements and changes to the specification without notice.  
Webtec behält sich das Recht vor, Verbesserungen oder Änderungen der Spezifikationen ohne Ankündigung vorzunehmen.  
Webtec se réserve le droit d'améliorer et de changer ses spécifications sans préavis.  
Webtec se reserva el derecho de realizar mejoras y cambios a las especificaciones sin previo aviso.

07/16

Version 6.0

FIRMG106-IN-ENG-2850.pdf

Designed and produced by Webtec - Entwickelt und hergestellt von Webtec  
Conçu et produit par Webtec - Diseñado y producido por Webtec